

Potentially fatal errors with GDH-PQQ [glucose dehydrogenase pyrroloquinoline quinone] glucose monitoring technology.

March 12, 2010

MedWatch Safety Alert. Silver Spring, MD: US Food and Drug Administration; August 13, 2009.

<https://psnet.ahrq.gov/issue/potentially-fatal-errors-gdh-pqq-glucose-dehydrogenase-pyrroloquinoline-quinone-glucose>

This public health notification raises awareness of the potential for falsely elevated blood glucose readings in patients using therapeutic products containing certain non-glucose sugars.