

Error tracking in a clinical biochemistry laboratory.

November 11, 2009

Szecsí PB, Ødum L. Error tracking in a clinical biochemistry laboratory. Clin Chem Lab Med. 2009;47(10). doi:10.1515/cclm.2009.272.

<https://psnet.ahrq.gov/issue/error-tracking-clinical-biochemistry-laboratory>

This study analyzed nearly 1200 [laboratory errors](#) through a systematic tracking approach and found that more than 80% were due to human error.