

Patient misidentifications caused by errors in standard barcode technology.

October 13, 2010

Snyder ML, Carter A, Jenkins K, et al. Patient misidentifications caused by errors in standard bar code technology. Clin Chem. 2010;56(10):1554-60. doi:10.1373/clinchem.2010.150094.

<https://psnet.ahrq.gov/issue/patient-misidentifications-caused-errors-standard-barcode-technology>

This study describes misidentification errors resulting from barcode–related [problems](#) with suboptimal printing of labels and inadequate scanner resolution requirements.