

Hospira Carpuject pre-filled cartridges—drug alert: products may contain more than the intended fill volume.

May 30, 2012

MedWatch Safety Alert. Silver Spring, MD: US Food and Drug Administration; May 23, 2012.

<https://psnet.ahrq.gov/issue/hospira-carpuject-pre-filled-cartridges-drug-alert-products-may-contain-more-intended-fill>

This announcement raises awareness of pre-filled medication cartridges that may be overfilled, thereby increasing the risk of overdose. The FDA recommends that practitioners confirm the dosage prior to dispensing and administering the medication.