

Video analysis of factors associated with response time to physiologic monitor alarms in a children's hospital.

August 3, 2017

Bonafide CP, Localio R, Holmes JH, et al. Video Analysis of Factors Associated With Response Time to Physiologic Monitor Alarms in a Children's Hospital. *JAMA Pediatr.* 2017;171(6):524-531.

doi:10.1001/jamapediatrics.2016.5123.

<https://psnet.ahrq.gov/issue/video-analysis-factors-associated-response-time-physiologic-monitor-alarms-childrens-hospital>

Bedside monitors alert nurses to [clinical deterioration](#). This prospective observational study examined nurse responses to bedside physiologic monitors. The mean response time was over 10 minutes. Less than 1% of alarms were actionable, underscoring the importance of addressing [alarm fatigue](#).